


Facility Requirements

Dispenser Dimensions: 48" wide x 39" deep x 78" high



Delivery Requirements:

Facility entrance must be ADA compliant. Entrance doorway and internal access doors that lead to the dispenser location, must meet the standard door dimensions of 36" W and 80" H. Please contact InstyMeds for unique delivery needs.

Location Requirements:

The dispenser should be located 4" from the back wall, have a 4" clearance above, and have a 12" clearance from any sidewall for field service access. Please contact InstyMeds for unique location circumstances.

Space in front of the dispenser should allow for wheelchair access.

To encourage use, place the dispenser in the waiting room area where it is highly visible to waiting patients and yet provides some level of privacy.

Note: The dispenser should not be installed next to food or beverage vending machines.

Electrical Requirements:

A duplex receptacle with 120v power needs to be located directly behind the dispenser about 18 inches above the floor. It is preferred that this duplex receptacle be wired as a 20-amp dedicated circuit.

Optional Power Upgrade: Emergency backup generator power will provide additional power reliability.

Internet Connection Requirements:

Two (2) Standard Ethernet jacks, installed in a RJ45 wall plate, need to be located directly behind the dispenser.

VPN: An internet router or security device capable of NAT (Network Address Translation) through a LAN-to-LAN VPN connection needs to be installed.

Telephone Connection Requirement:

A standard phone jack RJ11, located on the wall behind the dispenser, needs to be installed. The phone line should be set to "autodial" the InstyMeds Call Center at 952-653-2568.



Please contact InstyMeds for unique facility circumstances.

Call 1-866-467-8963

or email: info@instymeds.com

6501 City West Parkway • Eden Prairie, MN 55344

www.instymeds.com